

EU DECLARATION OF CONFORMITY

Manufacturer: **Mercator Medical (Thailand) Ltd.**
88/8 Moo.12 Tambon Kampaengphet,
Amphur Rattaphum, Songkhla 90180 Thailand

Declare under its sole responsibility that protective gloves:

Brand	Type	Sizes	Reference Numbers
MERCATOR ideall grip+	Nitrile, powder free, black, diamond textured, for single use	S (6-7) - XXL (10-11)	RP30026002-06_0001
MERCATOR ideall grip+	Nitrile, powder free, orange, diamond textured, for single use	S (6-7) - XXL (10-11)	RP30027002-06_0001

The products described above are in conformity with the provision of the PPE Regulation (EU) 2016/425 as a Category III product, Type B where such is the case, with standards EN ISO 374-1:2016+A1:2018, EN 374-2:2014, EN ISO 374-4:2019, EN ISO 374-5:2016, EN 16523-1:2015+A1:2018 and EN ISO 21420:2020.

The products are identical to Personal Protective equipment, which is the subject to the EU Type Examination (Module B) and Module C2 On-going Conformity under certificate No. 2777/14031-06/E00-00 issued by Notified Body:


SATRA Technology Europe Limited (Notified Body number 2777)
Bracetown Business Park, Clonee Dublin 15, Dublin, Ireland

This declaration is supported by the Quality System approval to ISO13485:2016/EN ISO13485:2016 and ISO 9001:2015 issued by:

SGS United Kingdom Ltd
Rossmore Business Park Ellesmere Port Cheshire CH65 3EN UK

Date and place of issue:
12.11.2021, Thailand

Signed on the behalf of the Manufacturer:


Mr. Praneet Inthajak
Senior QA/RA Manager

Mercator Medical (Thailand) Ltd.



Mercator Medical (Thailand) LTD.
88/8 Moo 12
Tambon Kampaengphet
Amphur Rattaphum
Songkhla 90180 Thailand

ph.: +66 74 584 222
fax.: +66 74 584 223
enquiry@mercator.co.th
www.mercator.co.th

Sales Office
Mercator Medical S.A.
ul. H. Modrzejewskiej 30
31-327 Kraków, Poland

ph.: +48 (12) 66 55 480
fax.: +48 (12) 66 55 415
medical@mercatormedical.eu
www.mercatormedical.eu